

ALR HEARING EXCERPT OF TESTIMONY

TESTIMONY OF TERRY ROBINSON

OCTOBER 12, 2005

1 TERRY ROBINSON,
2 having been duly sworn, was examined and testified as
3 follows:

4 DIRECT EXAMINATION

5 BY MR. BOYD:

6 Q. Okay. Terry, did you bring the maintenance
7 records and the test information logs pertaining to
8 instrument No. 68-13018?

9 A. Yes, I did.

10 Q. Thanks. And so the maintenance records
11 pertaining to this consist of five pages; is that right?

12 A. Yes, sir.

13 Q. Basically what you're showing during the
14 applicable time period is you are showing that you checked
15 for RFI, you checked for interferent, again you checked for
16 RFI. All these are showing really no problems as far as
17 interferent or the RFI; is that correct?

18 A. Yes, sir.

19 Q. But I noticed on March 31, 2005, Christi
20 detected an unstable reference. Could that be the chopper
21 motor?

22 A. No. Actually, what that is, is that's what
23 the instrument detects, but when it's reported on the
24 display of the instrument and what is reported as in the
25 test log, it's called an improper zero set. I believe it's

1 number -- there is a key in the back there of the test log
2 and it's No. 10 on the operational message list.

3 Q. Okay.

4 A. An improper zero set generally means that the
5 person that's delivering the sample did not wait until the
6 auto zero set was completely finished. And on the EN
7 instrument, if the person blows too soon, there's a period,
8 a very short period, in there in which you won't get an
9 improper sample, but you will get an improper zero set, if
10 it occurs at a very small window during the auto zero set.

11 Q. So when the instrument is actually taking air,
12 room air, in during one of the purges, I suppose, and you
13 below too early during a purge, it's detecting alcohol vapor
14 and they think it's being introduced from the room?

15 A. Part of -- part of your statement there is
16 correct. When it actually occurs is after the air blank and
17 prior to the subject samples and prior to the reference
18 sample there's another step called the auto zero. It's one
19 that some people refer to them as birds. I don't understand
20 why, but there's little greater-than symbols. Eight of them
21 will scroll across the display and that's the auto zero set.

22 And if a person -- if the operator allows

23 a person or in this case if the TS were to blow just a
24 little bit before that auto zero set was done, since during
25 that step the instrument is expecting not to see any alcohol

1 because it's setting to zero and it does see -- does see
2 alcohol or even detects a breath sample being delivered,
3 even if there is no alcohol, such as with the TS in that
4 particular little time space, that's what it says. It will
5 say unstable reference. It will then say improper zero set
6 and when it prints the test record out, it will say improper
7 zero set and that's what is included -- excuse me, that's
8 what is recorded on the test log as well.

9 Q. Was this done during a check by --

10 A. Yes.

11 Q. -- by Christi Swearingen? So she actually
12 blew into the machine while it was in this status of the
13 zero set?

14 A. I will say she probably blew a little bit too
15 soon --

16 Q. Okay.

17 A. -- during the auto zero set.

18 Q. But does that mean that Christi had alcohol on
19 her breath?

20 A. No. You can -- you get the auto zeroes --
21 when you get the improper zero set, it's just a physically
22 that you are blowing a little bit too soon during the auto
23 zero set. Now, you can have alcohol present. You may not
24 have alcohol present, such as that case there where Christi

25 blew. It's just a -- it's a matter of providing a specimen
1 when this is one when the instrument does not expect there
2 to be one being delivered.

3 Q. Okay. Just the detect there's something going
4 through the flow sensor?

5 A. Right.

6 Q. Okay. Okay. And then again we check with the
7 interferent RFI and we're finding no errors for the 30 days
8 prior to and 30 days following this test; is that correct?

9 A. That's correct.

10 MR. BOYD: Okay. I'll offer DX-1.

11 THE COURT: It's admitted.

12 MS. ANGELMAN: No objection.

13 THE COURT: Okay. No. 1 is admitted.

14 Q. (By Mr. Boyd) And these are the test
15 information logs for that applicable period; is that
16 correct?

17 A. Yes.

18 MR. BOYD: I'll call this DX-2. I'll
19 offer the test information logs.

20 MS. ANGELMAN: No objections.

21 THE COURT: Admitted.

22 Q. (By Mr. Boyd) Okay. And then -- now, getting
23 some activity during this applicable time period when I'm
24 seeing question marks for some of these tests, can you tell
25 me what that means?

1 A. That's -- if you look at the heading on that
2 column --

3 Q. Uh-huh.

4 A. -- it says it's an abbreviation for reference
5 result.

6 Q. Okay.

7 A. Now, if that particular situation that
8 occurred on that test occurred before the reference analysis
9 was conducted, it would show question marks simply because
10 there was no result obtained for that step on that test.

11 Q. Okay. So what does this mean? When you see
12 the question marks, does that mean there was no result
13 obtained for a reference sample?

14 A. I'm trying to see -- okay. If you look in the
15 final column on that test that you are looking at, you see
16 the No. 10.

17 Q. Yeah.

18 A. That's another improper zero set.

19 Q. Okay.

20 A. Okay. On that subject test, same scenario
21 which I have just described to you. Okay. Now, the reason
22 -- what I can tell you about that is that that instance on
23 that test occurred during the first sample delivery. The
24 reason I can say that is because if you go over to the next
25 column to the left under reference result, there are

1 question marks and that means that that step had not yet

2 been done by the instrument. That step that I'm referring
3 to is the reference analysis and that comes after the first
4 subject analysis.

5 Q. So somebody tried to blow during the air blank
6 or what?

7 A. No. If you blow during the air blank, you are
8 going to get a different indication called improper sample.

9 Q. Okay. So at what stage did this person
10 apparently blow?

11 A. That, again, same scenario, blew during the
12 auto zero set, just before it was finished.

13 Q. Okay.

14 A. Okay. You can -- if you blow as soon as the
15 first symbol, greater-than symbol, starts to scroll across,
16 you will get an improper sample. If it's almost finished
17 and there are eight little indicators that come up, about
18 one each half a second, and you don't let the words "please
19 blow into mouthpiece," you don't let that come up before the
20 subject blows, in other words, if you blow towards the end
21 of the auto zero set, that's what you get is an improper
22 zero set. And that's what that No. 10 means.

23 Q. I also noticed on 3-10, Mr. Pinkly (phonetic)
24 was doing an inspection test. He had these question marks
25 again.

1 A. Right.

2 Q. And this time you had message error reading 4?

3 A. Okay. I believe No. 4 is an interferent and

4 that's when we entered, intentionally introduced an acetone
5 sample into the sample chamber by blowing through a
6 simulator which has a water and acetone solution in it.
7 Now, we do that as -- we do that during the first sample
8 request. So, there again, the test will be invalidated, so
9 that the question marks represent that the reference
10 analysis was not conducted. That's why the question marks
11 are there.

12 Q. Okay. And, okay. Now, after the March events
13 where you had the spurious results from one of your --
14 again, this is -- okay. Now, this is -- of course, this
15 instrument is what you call the EN model machine, the
16 five-digit intoxilyzer; is that correct?

17 A. Yes, sir.

18 Q. Okay. Now, after the March events where you
19 had the spurious results, SWIFS crated up the two involved
20 machines in the false negative tests and shipped them back
21 to the Scientific Director or to CMI?

22 A. We hand carried them to Austin.

23 Q. Okay. And did you communicate to them what
24 the problems were and what you wanted them to check out?

25 A. We just told them what the problem was and we
1 left it up to them and their expertise to investigate it and
2 to see if they could come up with a reason or solution to
3 why they were doing it.

4 Q. And did they come up with that solution?

5 A. Yes, they did.

6 Q. And was that the memo from the Scientific
7 Director --

8 A. Yes, sir.

9 Q. -- to all the technical supervisors?

10 A. Yes, sir, that's correct.

11 Q. Okay. Did -- to your knowledge did the
12 Scientific Director ever communicate these problems to CMI?

13 A. Yes, sir, I believe they did, because after --
14 they kept the instruments in Austin for approximately a
15 month once they figured out what the problem was to evaluate
16 further. And then they were in communication at some point,
17 I don't know exactly when, but they shipped the instruments
18 from their laboratory to CMI sometime in April, I believe.

19 Q. Did CMI report back to the Scientific Director
20 with regard to their activities as far as checking out these
21 problems related to the EN models?

22 A. I believe what occurred was that the Office of
23 the Scientific Director reported the results that they
24 obtained as far as what the problem was that they found,
25 which is recorded in the letter that I believe you have a
1 copy of.

2 Q. Yeah.

3 A. And passed that information on to CMI as to
4 what they found. And I believe all CMI did was basically
5 duplicate what the Office of the Scientific Director had
6 done and just to satisfy for themselves that that could

7 occur. And between the two of them, they have gone back and
8 forth as far as recommendations about what to do to
9 alleviate that problem, so that it wouldn't occur in the
10 future.

11 Q. Did CMI make any recommendations with regard
12 to what could be done to the instruments to make them better
13 or not suffer from that problem?

14 A. Actually, I think the recommendation actually
15 was a combination of both CMI's engineers and the Office of
16 the Scientific Director about what they had -- about what
17 their plan was to do.

18 Q. What was their plan?

19 A. Their plan was to remove the small celluloid
20 valve that's in the instrument and place some larger
21 celluloid valves in them, which didn't directly have
22 anything to do with this issue. The more directly regarding
23 this issue --

24 [At this time the tape was turned
25 over by the Court.]

1 A. Not only on the EN, but on the 68 and in
2 operator's school, I actually somebody blew all the way so
3 there was 16. And when it does that, it blanks out and it
4 starts over again. I think I actually saw a guy do about 23
5 asterisks.

6 Q. Okay.

7 A. He was a triathlete.

8 Q. So it could just go on forever?

9 A. Yes. As long as the person continues to blow,
10 yes, sir.

11 Q. Okay. Now, either you or Christi had told me
12 that SWIFS was in the process of purchasing equipment to
13 calibrate the flow sensors on the ENs. Did you ever
14 purchase that additional equipment?

15 A. Yes, we have that.

16 Q. What did you get?

17 A. What did we get? A flow meter and a
18 compressor.

19 Q. Okay. And who did you purchase that from?

20 A. The compressor we got from Home Depot and the
21 flow meter we got from VWR Scientific.

22 Q. Okay. Now, they have plans or CMI had a plan
23 to put together a package of replacement parts to put into
24 these EN machines to alleviate this problem that you were
25 experiencing in March; is that correct?

1 A. Yes, sir.

2 Q. But did they ever put those packages together?

3 A. Not to my knowledge. It has not been done
4 yet.

5 Q. So you have been doing like ad hoc work on
6 these ENs, trying to get to accomplish a similar result?

7 A. No, sir. Through the recommendation of the
8 Department of Public Safety's Office of the Scientific
9 Director, we have checked all of our ENs with a vacuum to

10 see if we could get the valves to stick. If we couldn't get
11 them to stick, we just leave them alone. They're fine.
12 They won't -- it won't happen because the vacuum pressure
13 we're using is about ten times anything that a human being
14 could suck back into that instrument and cause that valve to
15 stick. So if it's not sticking, we just leave it. The only
16 one that we have actually changed is the one that was
17 involved in this test.

18 Q. The one that produced the zero result?

19 A. Right.

20 Q. Okay. Have you ever tried to stick the valves
21 on the aft end of the sample chamber?

22 A. The exhaust valve?

23 Q. Yes.

24 A. Yeah. You are basically doing the same thing
25 with -- you can do it with the vacuum pump, if you put your

1 finger over the valve, the one-way valve, in the back and
2 pump it up with the vacuum. You can do the same thing and
3 try to get it to stick.

4 Q. Can you get those to stick?

5 A. I have never been able to do that.

6 Q. Okay. So what would happen, theoretically, if
7 the -- if the one-way valves on the aft end of the sample
8 chamber, the exhaust valves, if they stuck?

9 A. It would basically close the system. Instead
10 of it being open, it would be closed.

11 Q. Could it pressurize the sample chamber?

12 A. Potentially, yes, sir.

13 Q. And if you pressurize the sample chamber, what
14 would you get as far as a subject test?

15 A. You could probably, because basically that's
16 what happened with the zero test, only it wasn't because of
17 that valve sticking, it was because of the one-way valve
18 sticking and that's in the block attached to the sample
19 chamber. Basically, the same thing would happen. It would
20 be possible to get a valid test, only in this case the
21 sample would reach the sample chamber and there potentially,
22 if the subject blew long enough, it could slightly elevate
23 the alcohol concentration because of the pressurization.

24 Q. Okay. Now, you were the Associate Editor of
25 the Alcohol Testing Alliance, is that correct, the Journal

1 of the Alcohol Testing Alliance, is that correct, back in
2 June of 1997?

3 A. Yes, sir. I have been a couple of times.

4 Q. Okay. And you were involved in the comparison
5 of the breath specimen to simultaneous blood draws; is that
6 correct?

7 A. I believe I peer reviewed that information,
8 yes, sir. I wasn't actually involved in that.

9 Q. So you weren't present during the testing?

10 A. I don't recall if that study was one that was
11 done at Texas A&M. I was involved in one that was done at
12 DPS.

13 Q. Okay. Were you -- did they establish a
14 criteria? When they did those simultaneous breath and blood
15 draws, did they establish a criteria for how many asterisks
16 they would try to obtain on the breath test, whether 1 or
17 23?

18 A. I don't recall.

19 Q. Okay. You don't know how they did that?

20 A. It was a long time ago. I don't recall.

21 Q. Okay. Now, do you know if the Scientific
22 Director has since conducted any experiments related to
23 conducting, trying out one asterisk sample for a valid test
24 versus 10 or 16 or 20 asterisk samples and comparing those
25 two samples of two valid breath tests conducted in close
1 proximity?

2 MS. ANGELMAN: I would object to the
3 relevance. I don't know how that's relevant to this.

4 THE COURT: I'll let it go, but you've
5 got one minute.

6 A. Since that study was done, I don't believe --
7 I may be mistaken, but I don't believe and I'm not -- I
8 don't think there have been any specific blood-to-breath
9 studies done by the Scientific Director.

10 Q. (By Mr. Boyd) Okay. I did not see any
11 criteria in the Journal of the Alcohol Testing Alliance
12 related to that June '97 issue as far as how many asterisks
13 -- of them recording a number of asterisks. Do you know

14 that they did or did not?

15 A. I doubt very seriously, if they did. But,
16 again, I would have to look at the research, but I doubt if
17 they did.

18 Q. Wouldn't that be a relevant test for you --
19 wouldn't you think that would be a relevant test to do to
20 compare one asterisk versus six or eight asterisks and
21 compare those to blood?

22 A. Well, you have got to remember all the
23 asterisks are showing you is how long someone blows.

24 Q. Yeah, but the length of the blow can increase
25 the --

1 A. That's incorrect. That's an incorrect
2 statement right there, because what it is that causes the
3 difference in the numbers is the volume of the sample
4 provided, not how long they blow. And there's no way
5 currently with this instrumentation to measure the volume of
6 the sample.

7 THE COURT: Okay, that's it.

8 [End of Excerpt of Testimony]

DEFENSE EXHIBIT NO. 1

Intoxilyzer maintenance records.

DEFENSE EXHIBIT NO. 2

Text information logs.

STATE OF TEXAS *
COUNTY OF DALLAS *

I, NANCY BREWER, Certified Court Reporter, do hereby certify that the foregoing pages constitute a full, true, and accurate transcript of the proceedings held in the above-captioned and numbered cause, as they were transcribed from audio tapes, to the best of my ability.

SUBSCRIBED to by me on this the 12th day of October, 2005.

NANCY BREWER, CSR, NO. 5759
My Commission Expires 12-31-2006